JUN - 6 2003

510(k) Summary
for the 1 mL Bak'Snap DuoPro™ Safety Syringe (DuoProSS™)
(per 21CFR807.92)

1. SPONSOR

M.K. Meditech Co., Ltd.
Suite 702, 7th Floor
No. 5, Chingdau E. Rd.
Taipei, Taiwan 100.

Contact Person: I-Ming Shih
Telephone: 886-2-23930855

Date Prepared: June 5, 2003

2. DEVICE NAME

Proprietary Name:	Bak'Snap and Safety'Tray Models of the DuoPro™ Safety Syringe (DuoProSS™)
Common/Usual Name:	Hypodermic Syringe (with needle) (tuberculin-type)
Classification Name:	Piston syringe Hypodermic single lumen needle

3. PREDICATE DEVICES

- DuoPro™ Safety Syringe (DuoProSS™) (K020623 & K022806)
- SECUREGARD® Retractable Safety Syringe (K012121)

4. DEVICE DESCRIPTION

The 1 mL Bak'Snap DuoPro™ Safety Syringe (DuoProSS™) is a sterile, single-use, disposable and non-reusable, manual, retractable safety syringe provided with permanently attached needle. The Safety'Tray Safety Allergy Syringe Tray consists of 25 Bak'Snap syringes within a single sterile barrier.

5. INTENDED USE

The 1 mL Bak'Snap DuoPro™ Safety Syringe (DuoProSS™) is a sterile, single-use, disposable and non-reusable, manual, retractable safety syringe which is intended for injection of fluids into the body, while reducing the risk of sharps injuries and the potential for syringe reuse.

6. TECHNOLOGICAL CHARACTERISTICS AND SUBSTANTIAL EQUIVALENCE

M.K. Meditech Co., Ltd., makes a claim of substantial equivalence of the 1 mL Bak'Snap DuoProSS™ to the cited predicates based on similarities in intended use, design, and technological and operational characteristics. They are indicated for injecting fluids into the body, while helping to reduce the risk of sharps injuries. Both the 1 mL Bak'Snap DuoProSS™ and the SECUREGARD® syringe have permanently attached, single-lumen hypodermic needles.

All syringes are provided sterile, single-use, and disposable. All syringes have two-part plungers. The distal part holds the hypodermic needle and the proximal part has a projection spike that mates with the distal part, thereby locking the needle to the plunger. Both syringes require the user to manually retract the needle-plunger into the syringe barrel, break off the plunger rod, and discard the pieces. M.K. Meditech Co., Ltd., believes that the differences between the 1 mL Bak'Snap DuoPro™ Safety Syringe and cited predicate devices are minor and they raise no new issues of safety or effectiveness.

7. TESTING

Verification and validation testing presented in this premarket notification includes testing to demonstrate conformance to standards, testing according to FDA guidance, biocompatibility per ISO 10993, and sterilization re-validation.

DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

JUN - 6 2003

M.K. Meditech Company Limited
C/O Ms. Rosina Robinson
Medical Device Consultants, Incorporated
49 Plain Street
North Attleboro, Massachusetts 02760

Re: K031594
Trade/Device Name: Bak' Snap DuoPro™ Safety Syringe (DuoProSS™)
Regulation Number: 21 CFR 880.5860
Regulation Name: Piston Syringe
Regulatory Class: II
Product Code: MEG
Dated: May 20, 2003
Received: May 21, 2003

Dear Ms. Robinson:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (301) 594-4618. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address http://www.fda.gov/cdrh/dsma/dsmamain.html

Sincerely yours,

Susan Runner

Susan Runner, DDS, MA
Interim Director
Division of Anesthesiology, General Hospital,
Infection Control and Dental Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

510(k) Number (if known): K031594

Device Name: M.K. Meditech Co., Ltd.,
1 mL Bak'Snap DuoPro™ Safety Syringe (DuoProSS™)

Indications For Use:

The 1 mL Bak'Snap DuoPro™ Safety Syringe (DuoProSS™) is a sterile, single-use, disposable and non-reusable, manual, retractable safety syringe intended for injection of fluids into the body, while reducing the risk of sharps injuries and the potential for syringe reuse.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NECESSARY)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Rebecca Cucante

(Division Sign-Off)
Division of Anesthesiology, General Hospital,
Infection Control, Dental Devices

510(k) Number: K031594

Prescription Use ✓ OR Over-The-Counter Use ____
(Per 21 CFR 801.109)

(Optional Format 1-2-96)